

OVERVIEW Z-SCORES

Satisfactory ($-2 \leq Z\text{-score} \leq 2$)
 Need attention ($-3 \leq Z\text{-score} < -2$ or $2 < Z\text{-score} \leq 3$)
 Unsatisfactory ($Z\text{-score} < -3$ or > 3)

Module	Parameter	Sample	Test System	Total	Assay	Method
Coagulation Factor - I	Factor VIII (clotting activity)	19.70	1	-1.08	-1.09	-0.55
		19.71	1	1.29	1.27	2.80
	Factor IX (clotting)	19.70	1	-0.19	-0.19	0.18
		19.71	1	1.17	1.15	2.84
	Factor XI	19.70	1	0.29	0.28	0.61
		19.71	1	1.88	1.88	3.33
Coagulation Factor - II	Factor XII	19.70	1	0.26	0.25	-0.36
		19.71	1	1.50	1.49	1.54
	Factor II	19.72	1	3.16	3.16	4.19
		19.73	1	-1.01	-1.01	0.28
Coagulation Factor - II	Factor V	19.72	1	2.85	2.85	4.10
		19.73	1	-0.83	-0.83	-0.21
	Factor VII	19.72	1	2.11	2.11	3.90
		19.73	1	40.61	40.61	58.55
	Factor X	19.72	1	2.59	2.59	4.33
		19.73	1	-1.48	-1.48	-1.25
Von Willebrand Factor Parameters	VWF antigen	19.74	1	1.04	0.98	0.26
	VWF activity	19.74	1	-0.14	-0.01	-0.01
	VWF Collagen Binding Assay	19.74	1	-0.42	-0.06	0.40
	VWF Factor VIII	19.74	1	0.73	0.69	
Factor XIII	Factor XIII activity (quantitative)	19.79	1	0.18	0.18	0.23
		19.80	1	-0.58	-0.58	-0.82
Fibrinolysis - I	Plasminogen activity	19.81	1	0.67	0.67	0.15
		19.82	1	-0.82	-0.82	0.37
	Plasmin Inhibitor	19.81	1	0.69	0.69	0.47
		19.82	1	1.24	1.24	1.20
Fibrinolysis - II	PAI-1 antigen	19.83	1	1.59	1.59	
		19.84	1	1.65	1.65	
	t-PA antigen	19.83	1	-0.63	-0.63	
		19.84	1	-1.12	-1.12	