

OVERVIEW Z-SCORES

■ UV: Satisfactory ($-2 \leq Z\text{-score} \leq 2$)
 ■ UV: Need attention ($-3 < Z\text{-score} < -2$ or $2 < Z\text{-score} < 3$)
 ■ UV: Unsatisfactory ($Z\text{-score} \leq -3$ or ≥ 3)
 ■ BV: Unacceptable

Module	Parameter	Sample	Test System	Total	Assay	Method	Bivariate Analysis
Coagulation Factor - I	Factor XI	23.45	1	-0.21	-0.21	0.46	
		23.46	1	-0.74	-0.73	-0.76	
Coagulation Factor - I	Factor XII	23.45	1	0.90	0.90	1.35	
		23.46	1	1.02	1.03	1.08	
Coagulation Factor - II	Factor II	23.47	1	0.04	0.04	0.39	
		23.48	1	-0.76	-0.76	-0.23	
Coagulation Factor - II	Factor V	23.47	1	0.72	0.72	1.87	
		23.48	1	0.60	0.60	1.90	
Coagulation Factor - II	Factor VII	23.47	1	-0.43	-0.43	0.30	
		23.48	1	-0.29	-0.29	0.31	
Coagulation Factor - II	Factor X	23.47	1	0.12	0.13	0.50	
		23.48	1	-0.69	-0.69	-0.24	
Von Willebrand Factor Par	VWF antigen	23.49	1	1.06	1.00	0.12	
		23.50	1	0.57	0.52	-0.67	
Von Willebrand Factor Par	VWF Collagen Binding Assay	23.49	1	1.55	0.74	0.07	
		23.50	1	0.53	1.07	1.09	
Von Willebrand Factor Par	VWF Factor VIII	23.49	1	0.95	0.98		
		23.50	1	1.73	1.75		
Factor XIII	Factor XIII activity (quantitative)	23.53	1	-0.39	-0.40	-0.70	
		23.54	1	0.19	0.18	-0.06	
Fibrinolysis - I	Plasminogen activity	23.55	1	-2.09	-2.09	-0.88	
		23.56	1	-0.58	-0.58	0.31	
Fibrinolysis - I	Plasmin Inhibitor	23.55	1	0.61	0.61	1.16	
		23.56	1	0.83	0.83	0.88	
Fibrinolysis - II	PAI-1 antigen	23.57	1	-0.47	-0.47		
		23.58	1	0.96	0.96		
Fibrinolysis - II	t-PA antigen	23.57	1	-1.16	-1.16		
		23.58	1	-0.59	-0.59		
Von Willebrand Factor Par	VWF activity	23.49	1	0.62	0.19	0.02	
		23.50	1	0.77	0.37	0.26	
Coagulation Factor - I	Factor VIII (clotting activity)	23.45	1	0.88	0.89	2.10	
		23.46	1	1.45	1.44	2.34	
Coagulation Factor - I	Factor IX (clotting)	23.45	1	-0.36	-0.36	0.38	
		23.46	1	0.13	0.13	0.64	