

OVERVIEW Z-SCORES

■ UV: Satisfactory ($-2 \leq Z\text{-score} \leq 2$)
 ■ UV: Need attention ($-3 < Z\text{-score} < -2$ or $2 < Z\text{-score} < 3$)
 ■ UV: Unsatisfactory ($Z\text{-score} \leq -3$ or ≥ 3)
 ■ BV: Unacceptable

Module	Parameter	Sample	Test System	Total	Assay	Method	Bivariate Analysis
Coagulation Factor - I	Factor XI	23.146	1	-0.30	-0.29	-0.38	
		23.147	1	-0.76	-0.76	-0.49	
Coagulation Factor - I	Factor XII	23.146	1	-1.30	-1.29	-1.40	
		23.147	1	-0.93	-0.93	-0.89	
Coagulation Factor - II	Factor II	23.148	1	-0.40	-0.40	-0.18	
		23.149	1	-1.07	-1.07	-0.67	
Coagulation Factor - II	Factor V	23.148	1	-0.82	-0.82	-0.48	
		23.149	1	-1.20	-1.20	-0.93	
Coagulation Factor - II	Factor VII	23.148	1	-0.42	-0.42	1.16	
		23.149	1	-0.32	-0.32	0.30	
Coagulation Factor - II	Factor X	23.148	1	1.12	1.12	0.85	
		23.149	1	0.24	0.24	0.74	
Von Willebrand Factor Par	VWF antigen	23.150	1	1.31	1.20	0.59	
		23.151	1	1.45	1.47	0.54	
Von Willebrand Factor Par	VWF Collagen Binding Assay	23.150	1	3.21	2.22	1.29	
		23.151	1	0.92	0.89	0.29	
Von Willebrand Factor Par	VWF Factor VIII	23.150	1	-0.77	-0.86		
		23.151	1	-0.97	-1.00		
Factor XIII	Factor XIII activity (quantitative)	23.154	1	-0.58	-0.61	-0.90	
		23.155	1	-1.24	-1.25	-1.53	
Fibrinolysis - I	Plasminogen activity	23.156	1	1.12	1.12	0.92	
		23.157	1	-1.99	-1.99	-0.83	
Fibrinolysis - I	Plasmin Inhibitor	23.156	1	-0.41	-0.41	-1.00	
		23.157	1	-0.45	-0.45	-0.19	
Fibrinolysis - II	PAI-1 antigen	23.158	1	0.59	0.59		
		23.159	1	-0.72	-0.72	-1.12	
Fibrinolysis - II	t-PA antigen	23.158	1	-0.83	-0.83		
		23.159	1	-0.82	-0.82		
Von Willebrand Factor Par	VWF activity	23.150	1	0.77	0.46	0.37	
		23.151	1	-0.30	-0.38	-0.43	
Coagulation Factor - I	Factor VIII (clotting activity)	23.146	1	-0.58	-0.58	-0.34	
		23.147	1	-0.10	-0.11	0.37	
Coagulation Factor - I	Factor IX (clotting)	23.146	1	1.01	1.00	1.34	
		23.147	1	-0.60	-0.60	-0.14	