



DEKS

Rigshospitalet - Glostrup
Valdemar Hansens Vej 1-23
Opgang 8, 1.
2600 Glostrup
Denmark
☎ +45 3863 4400
www.deks.dk

Test period:
11.10.2021 – 22.10.2021
Last reporting day: 25.10.2021

Scheme coordinator:
Sanne Schou
☎ +45 38634405
sanne.schou@deks.dk
Gitte M. Henriksen
☎ +45 38634401
gitte.henriksen@deks.dk

This report contains:
• This cover letter
• Summary report
• Individual results

Approved by GMH
08.11.2021

Next reporting period starts
29.11.2021



copyright © DEKS

Methylmalonic Acid 3033 DK EQA rapport 4/ 2021

Number of participants

There are 44 laboratories registered for this scheme. Most of the participants are from the Scandinavian countries and 7 laboratories are from other countries. DEKS has received 37 results for this report.

Sample material

Sample no. 07_2021 "Hgj" is a pool of serum. This pool has not been measured before.

Sample no. 08_2021 "lqt" is a pool of serum. This pool has been measured once before.

Statistics

In the graphical part you will find your own results and deviation (dev.) from the mean (M).

Target value

The target value is calculated as the mean of all.

Acceptance

The acceptance interval is $\pm 21\%$. The acceptance interval is calculated based on biological variation¹.

Results and comments

Two outliers was found in this round. Outliers are defined as $>SD*3,6$ from the overall mean.

Kind regards

Gitte Henriksen og Sanne Schou

1) Moeller J, Rasmussen K, Christensen L. External quality assessment of methylmalonic acid and total homocysteine. Clin Chem 1999;45:9:1536-152.

